

AURELIA® Sonic®

Examination Gloves



9377B5	(XS)	ø 1300	10 x 100
9377B6	(S)	ø 1300	10 x 100
9377B7	(M)	ø 1300	10 x 100
9377B8	(L)	ø 1300	10 x 100
9377B9	(XL)	ø 1300	10 x 100

Nitrile

Powder Free

Cobalt Blue

3.2+/- 0.2 g

Finger-Textured

Quality Cobalt Blue Nitrile examination gloves that offer increased tensile strength and protection over Latex and Vinyl.

Incredible Ultra-Thin Nitrile examination gloves specifically designed to offer increased comfort, flexibility & sensitivity.

- 2.2mil Powder-Free Nitrile
- Non-Sterile
- Beaded cuff
- Ambidextrous
- Latex Free
- AQL 1.5

Sizes & Dimensions



XS 80 ± 10 mm S 95 ± 10 mm М 110 ± 10 mm L XL

 $70 \pm 10 \text{ mm}$ min. 240 mm ≥ 110 mm

Single wall thickness



① min. 0.11 mm

② min. 0.05 mm

③ min. 0.05 mm

Physical properties

	Unaged	Aged
Tensile Strength	min. 14.0 MPa	min. 14.0 MPa
Elongation	min. 500 %	min. 400 %
Strength at Break	min. 6N	min. 6N

Aurelia Gloves conforms and complies with:

- The general safety and performance requirements of FDA Medical Device Regulation for Class 1 medical devices (21 CFR 880.6250) and complies with all general controls (section 513(a)(1)(A) of the Federal Food, Drug, and Cosmetic Act (Act); 21 U.S.C. 360c(a)(1)(A))
- The standard specification requirements for ASTM D6319- Freedom From Holes, Physical Dimensions Test, Physical Requirement Test and Packaging
- The standard specification requirements for ASTM D6978-05 Standard Practice for Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs
- EEC regulations concerning the conformity of materials and products that are allowed to come into contact with food. In accordance with Regulation EEC 1935/2004, Regulation EC 10/2011 & Regulation (EC) No 2023/2006.
- The gloves are certified manufactured according to ISO 9001:2015 and ISO 13485:2016 Quality Management Systems